

BELL SYSTEM PRACTICES
Central Office Maintenance
Description of Testing
Equipment and Tools
Toll Test Room Operation
Description and Operating
Principles of Systems and
Equipment
Radio Systems
General

ADDENDUM A702.636
ADDENDUM E40.571
ADDENDUM R70.550
Issue 1, October, 1955
AT&TCo Standard

72A FREQUENCY METER

(J64072A)

1. GENERAL

1.011 This addendum, Issue 1 modifies Issue 1 of the section.

1.021 The purpose of this addendum is to make several minor numerical corrections in the figures and text of the section.

2. DESCRIPTION

2.22 Interchange "horizontal" and "vertical" in the 4th and 5th lines.

4. OPERATION

(C) Calibration

Change Fig. 10H on Page 10 to indicate 3000 cycles instead of 300 cycles.

(F) Frequency Measurements Outside Normal Operating Range

Change Figs. 13A and B on Page 13 to indicate for Fig. A a range of 4000 to 40,000 cycles and for Fig. B a range of 10 to 100 cycles.

4.11

(3) Interchange the references to Figs. 13A and 13B to be consistent with the above changes in these figures.